

Eppendorf Certificate


Certificate of Purity – PCR clean

This package contains a high-quality consumable manufactured under the PCR clean Eppendorf Purity Standard.

The Eppendorf PCR clean consumables are produced in an ISO class 8 clean room environment according to ISO 14644-1. For this product Eppendorf certifies the following[*]:

Free of detectable

- Human DNA
- DNase
- RNase
- PCR inhibitors



[*] Filtertips are additionally sterile & free of pyrogens, UVettes are free of protein.

These parameters are continuously monitored by an independent certified laboratory. Eppendorf guarantees the conformity within the following limits:

Human DNA	< 2 pg
DNase	< 1.0×10^{-6} Kunitz units
RNase	< 1.0×10^{-9} Kunitz units
PCR inhibitors	less than 10 targets amplifiable

Quality control and subsequent certification is done by an independent laboratory. Lot-related certificates are available on request or on the internet at www.eppendorf.com/certificates.

The certification comprises the following tests:

Human DNA Contamination Test

A PCR master mix is prepared using the QuantiTect® SYBR® Green PCR Kit (QIAGEN®) and primer for the detection of human DNA. The primers amplify a 294 bp fragment present in more than 1×10^5 copies per human cell. The master mix (20 µl) is added to 5 positive control vessels containing known amounts of human DNA (32, 16, 8, 4 and 2 pg in 5 µL H₂O) plus a negative control (10 µL DNA-free H₂O).

15 samples are rinsed one after another with DNA-free water. 10 µL of this solution are added to 20 µL master mix. PCR is done for 30 cycles.

The emittance of SYBR Green-induced fluorescence is detected in samples and controls. For the samples to pass certification, no fluorescence must be found.

Eppendorf Certificate

DNase Test

15 samples are rinsed one after another with DNA-free water. 17 µL of these solutions are mixed with 3 µL DNase-buffer containing 100 bp-DNA-ladder in a DNase-free tube. A positive control is spiked with DNase, a negative control contains DNA-free water. All tubes are incubated for 24 h at 37 °C.

The DNA is analyzed by agarose-gelelectrophoresis. DNase contamination is indicated by degradation of the DNA ladder. For samples to pass certification, the relative intensities of the DNA pattern of the samples must correspond to the negative control.

RNase Test

15 samples are rinsed one after another with RNA-free water. 17 µL of these solutions are mixed with 3 µL RNase-buffer containing 100 bp-RNA-ladder in a RNase-free tube. A positive control is spiked with RNase, a negative control contains RNA-free water. All vessels are incubated for 24 h at 37 °C.

The RNA is analyzed by agarose-gelelectrophoresis. RNase contamination is indicated by degradation of the RNA ladder. For samples to pass certification, the relative intensities of the RNA pattern of the samples must correspond to the negative control.

PCR Inhibitor Test

A PCR master mix is prepared using the QuantiTect SYBR Green PCR Kit (QIAGEN®), primer for the detection of human DNA and 16 pg human DNA. The primers amplify a 294 bp fragment present in more than 10⁵ copies per human cell.

15 samples are rinsed one after another with DNA-free water. 10 µL of this solution are added to 15 µL master mix plus 16 pg human DNA. PCR is done for 30 cycles.

The emittance of SYBR Green-induced fluorescence is detected in samples and controls. For the samples to pass certification, the CT values of the samples are compared with the positive control (containing 16 pg human DNA). The difference of the CT-value between the samples and the control must be in range of +/- 2 cycles.

October, 2019

Page 2 of 2

0012557.051-08



Joana Tziolis
Expert Liquid Handling
LH-Consumables



Dr. Birgit Schreiber
Vice President Quality Management &
Regulatory Affairs

Your local distributor: www.eppendorf.com/contact
Eppendorf AG · Barkhausenweg 1 · 22339 Hamburg · Germany
E-mail: eppendorf@eppendorf.com

ISO 9001
Certified

ISO 13485
Certified

ISO 14001
Certified

QIAGEN® and QuantiTect® are registered trademarks of Qiagen GmbH, Germany. SYBR® is a registered trademark of molecular Probes, Inc., USA. Eppendorf® and the Eppendorf Brand Design are registered trademarks of Eppendorf AG, Germany.
All rights reserved, incl. graphics and images. Copyright ©2019 by Eppendorf AG.

www.eppendorf.com